



6390 Joyce Drive
#100
Golden, CO 80403

PHONE 303-940-0033
FAX 303-940-0043
www.wibby.com

**Instructions for:
WP Cyanide - Amenable to Chlorination (cat # PT-CNA-WP)**

STANDARD DESCRIPTION

- The WP Cyanide - Amenable to Chlorination standard is provided in a vial that contains approximately 21 mL of concentrate.
- The WP Cyanide - Amenable to Chlorination standard is unpreserved.
- Store the standard in the unopened vial at room temperature (~ 20 - 25°C).
- Avoid exposure of the vial to direct sunlight.

ADDITIONAL INFORMATION

- The standard has been provided as a concentrate that must be diluted prior to analysis.
- The WP Cyanide - Amenable to Chlorination standard is manufactured using simple cyanide (e.g., free, dissociable, amenable to chlorination). This standard is not compliant with NELAC/EPA criteria for a WP Cyanide standard. If you need a cyanide standard that is NELAC/EPA compliant (e.g. is manufactured from complex cyanide), order Wibby Environmental's WP Cyanide standard (catalog # PT-CN-WP).

STANDARD PREPARATION, ANALYSIS and STORAGE

1. For best results, the PT standards should be stored at room temperature (~ 20 - 25°C).
2. Add approximately 900 mL of ASTM Type 1 water to a 1000 mL class A volumetric flask.
3. Transfer exactly 10.0 mL of the PT standard concentrate to the flask using a class A volumetric pipette.
4. Add sufficient NaOH as a preservative to bring the pH of the standard to a pH >12.
5. Bring the flask to volume with ASTM Type 1 water.
6. Mix the solution by inverting the volumetric flask a minimum of three times.
7. The standard is now ready for preparation and analysis per the selected method(s).
8. The sample should be analyzed as soon as possible after dilution,
9. Report all results in mg/L per the reporting instructions contained in this booklet.
10. Store the diluted standard and any remaining concentrate at room temperature (~ 20 - 25°C).

CONCENTRATION RANGE and PTRL

- After preparation per these instructions, the standard will contain the analytes in the following table at a certified Concentration within the ranges shown.
- The NELAC Proficiency Testing Reporting Limit (PTRL) is provided as guidance when analyzing NELAC PT standards. At a minimum, the laboratory should use a method that is sensitive enough to generate quantitative results at the PTRL shown.

Analyte	Units	Concentration Range	PTRL
Cyanide - Amenable to Chlorination	mg/L	0.100 - 0.500	NA