

Instructions for: **WP Formaldehyde (Cat # PT-FORM-WP)**

SCOPE AND APPLICATION

Wibby Environmental's WP Formaldehyde Proficiency Testing Standard is designed to be used with promulgated EPA methods as well as applicable methods from consensus organizations such as AWWA and ASTM. You should be familiar with the method(s) before analyzing the standard. If you have any questions about the use of these standards, please contact Wibby Environmental Customer Service at 1-866-WibbyPT (866-942-2978).

SAMPLE PREPARATION AND ANALYSIS

1. For best results, the PT standard should be refrigerated and then brought to room temperature (near 20°C) when used.
2. Add approximately 990 mL of organic free deionized water to a 1000 mL Class A volumetric flask.
3. Carefully open the ampule by snapping off the top at the narrow part of the neck.
4. Transfer exactly 1.0 mL of the PT standard concentrate to the flask using a gas-tight syringe and delivering the aliquot below the surface of the water.
5. Bring the volumetric flask to volume with organic free deionized water.
6. Mix the solution by inverting the volumetric flask a minimum of three times.
7. The standard is now ready for preparation and analysis per the selected method(s). We recommend the standard be analyzed as soon as possible after dilution.

REPORTING RESULTS

1. Report results to three significant figures.
2. Report your results on line at www.wibby.com. Click on the "Online Data Entry" link or the "PT Manage" link.
3. You may also report your results using the Data Reporting Sheets enclosed with your standards. FAX your results to Wibby Environmental at 1-866-283-0269 or mail the results to Wibby Environmental, 6390 Joyce Drive, #100, Golden, CO, 80403.
4. Wibby Environmental must receive all results prior to the study closing date shown on the Data Reporting Sheets.

SAFETY

The standards are designed for use by laboratory professionals who are familiar with handling environmental reference materials as well as hazardous materials. If you have any questions about the safe handling of these standards or require a MSDS sheet, please contact Wibby Environmental at 1-866-WibbyPT (866-942-2978).

QUESTIONS?

If you have any questions regarding these standards, please call Wibby Environmental at 1-866-WibbyPT (866-942-2978).

6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.wibby.com

Water Pollution Proficiency Testing Concentration Ranges and PTRLs

Definitions:

PTRL

NELAC Proficiency Testing Reporting Limits (PTRLs) are provided as guidance to laboratories analyzing NELAC PT samples. At a minimum, the laboratory should use a method that is sensitive enough to generate quantitative results at the PTRLs shown. (REF: NELAC PT FOT Tables)

NA

Not Applicable (NA) has been applied to analytes where a PTRL is not applicable and to state specific analytes that have not had a PTRL determined by the applicable accrediting agency.

Formaldehyde (PT-FORM-WP)

NELAC Code	Analyte	Units	Concentration Range	PTRL
Additional State Specific Analytes				
4815	Formaldehyde	µg/L	10.0 - 200	NA