

Instructions for: WP Oil and Grease (cat # PT-OGR-WP)

STANDARD DESCRIPTION

- The WP Oil and Grease standard is provided as a concentrate that must be diluted prior to extraction and analysis.
- The WP Oil and Grease standard is preserved with approximately 0.1 % hydrochloric acid.
- Store the standard in the unopened vial at room temperature (~ 20 - 25°C).

ADDITIONAL INFORMATION

- The standard has been provided as a concentrate that must be diluted prior to analysis. We also offer a whole volume ready-to-use WP Oil and Grease standard (catalog #PT-OGR1L-WP). If you would like to order the 1 L version of this standard, please call customer service at 1-866-WibbyPT (866-942-2978).
- Oil and grease also has a tendency to adhere to the vial and cap. After transferring the concentrate to your extraction vessel or apparatus, rinse the standard bottle and cap a minimum of two times with the extraction solvent. The solvent used for rinsing the container should be combined with that used for extraction of the sample.

STANDARD PREPARATION, ANALYSIS and STORAGE

1. For best results, the Oil and Grease concentrate should be stored near room temperature (~20°C).
2. Add approximately 750 mL of ASTM Type 1 water to a 1000 mL class A volumetric flask or to a tared sample bottle.
3. Shake the Oil and Grease concentrate vial thoroughly to mix the sample.
4. Quickly and carefully transfer the entire contents of the vial containing the Oil and Grease concentrate to the volumetric flask or sample bottle.
5. Bring the sample to a final volume of one liter with ASTM Type 1 water.
6. The standard is now ready for preparation and analysis per the selected method(s).
7. The sample should be analyzed as soon as possible after dilution,
8. Report all results in mg/L per the reporting instructions contained in this booklet.
9. Store the diluted standard and any remaining concentrate at room temperature (~ 20 - 25°C).



6390 Joyce Drive
#100
Golden, CO 80403

PHONE 303-940-0033
FAX 303-940-0043
www.wibby.com

**Instructions for:
WP Oil and Grease (cat # PT-OGR-WP) - Continued**

CONCENTRATION RANGE and PTRL

- After preparation per these instructions, the standard will contain Oil and Grease at a certified Concentration within the range shown.
- The NELAC Proficiency Testing Reporting Limit (PTRL) is provided as guidance when analyzing NELAC PT standards. At a minimum, the laboratory should use a method that is sensitive enough to generate quantitative results at the PTRL shown.

Analyte	Units	Concentration Range	PTRL
Oil and Grease (n-hexane extractable material)	mg/L	20.0 - 100	8.80