



6390 Joyce Drive  
#100  
Golden, CO 80403

PHONE 303-940-0033  
FAX 303-940-0043  
www.wibby.com

## Instructions for: **WP Phenolics (cat # PT-PHEN-WP)**

### **STANDARD DESCRIPTION**

- The WP Phenolics standard is provided in a flame sealed ampule that contains approximately 2.1 mL of concentrate.
- The WP Phenolics standard is packaged in acetone.
- Store the standard in the unopened ampule refrigerated (at ~4°C).

### **ADDITIONAL INFORMATION**

- The standard has been provided as a concentrate that must be diluted prior to analysis.
- The WP Phenolics standard contains four sources of phenol (phenol, 2-chlorophenol, 2,4-dinitrophenol, 2,4-dichlorophenol) and per NELAC/EPA requirements is designed to be analyzed by distillation-4-AAP methods.

### **STANDARD PREPARATION, ANALYSIS and STORAGE**

1. For best results, the PT standard should be stored refrigerated and then brought to room temperature (near 20°C) when used.
2. Add approximately 990 mL of ASTM Type 1 water to a 1000 mL class A volumetric flask.
3. Carefully open the ampule by snapping off the top at the narrow part of the neck.
4. Transfer exactly 1.00 mL of the PT standard concentrate to the flask using a gas tight calibrated syringe.
5. Bring the flask to volume with ASTM Type 1 water.
6. Mix the solution by inverting the volumetric flask a minimum of three times.
7. The standard is now ready for preparation and analysis per your routine method(s).
8. The sample should be analyzed as soon as possible after dilution,
9. Report all results in mg/L per the reporting instructions contained in this booklet.
10. Store the diluted standard and any remaining concentrate refrigerated (at ~4°C).

### **CONCENTRATION RANGE and PTRL**

- After preparation per these instructions, the standard will contain the analytes in the following table at a certified Concentration within the ranges shown.
- The NELAC Proficiency Testing Reporting Limit (PTRL) is provided as guidance when analyzing NELAC PT standards. At a minimum, the laboratory should use a method that is sensitive enough to generate quantitative results at the PTRL shown.

Analyte	Units	Concentration Range	PTRL
Total Phenolics (4AAP)	mg/L	0.0600 – 5.00	0.010