



6390 Joyce Drive  
#100  
Golden, CO 80403

PHONE 303-940-0033  
FAX 303-940-0043  
www.wibby.com

## Instructions for: **WP Solids (cat # PT-SOL-WP)**

### **SCOPE AND APPLICATION**

Wibby Environmental's WP Solids Proficiency Testing Standard is designed to be used for US EPA Methods given in "Methods for Chemical Analysis of Water and Wastes" or methods given in "Standard Methods for the Examination of Water and Wastewater". If you have any questions about the use of these standards, please contact Wibby Environmental Customer Service at 1-866-WibbyPT (866-942-2978).

### **SAMPLE PREPARATION AND ANALYSIS**

1. For best results, the PT standard should be stored near room temperature (~ 20°C).
2. The standard is ready for preparation and analysis per your routine method(s) as received. No dilution is required.

**NOTE:** The Solids standard contains suspended solids. To ensure a uniform subsample, shake the bottle vigorously prior to removing an aliquot for analysis.

### **REPORTING RESULTS**

1. Report results to three significant figures.
2. Report your results on line at [www.wibby.com](http://www.wibby.com). Click on the "Online Data Entry" link or the "PT Manage" link.
3. You may also report your results using the Data Reporting Sheets enclosed with your standards. FAX your results to Wibby Environmental at 1-866-283-0269 or mail the results to Wibby Environmental, 6390 Joyce Drive, #100, Golden, CO, 80403.
4. Wibby Environmental must receive all results prior to the study closing date shown on the Data Reporting Sheets.

### **SAFETY**

The standards are designed for use by laboratory professionals who are familiar with handling environmental reference materials as well as hazardous materials. If you have any questions about the safe handling of these standards or require a Material Safety Data Sheet (MSDS,) please contact Wibby Environmental at 1-866-WibbyPT (866-942-2978).

### **QUESTIONS?**

If you have any questions regarding these standards or reporting requirements, please call Wibby Environmental at 1-866-WibbyPT (866-942-2978).

### **CONCENTRATION RANGE and PTRL**

1. The standard will contain the analytes listed below at a certified Concentration within the ranges shown.
2. The NELAC Proficiency Testing Reporting Limit (PTRL) is provided as guidance when analyzing NELAC PT standards. At a minimum, the laboratory should use a method that is sensitive enough to generate quantitative results at the PTRL shown.

Analyte	Conc Range	PTRL
Total Dissolved Solids (TDS)	140 – 650	98.0
Non-Filterable Solids (TSS)	23.0 – 100	14.0
Total Solids	140 - 675	106

1. Units for all analytes are mg/L.