

**Instructions for:
WP Volatile Solids (cat # PT-VSOL-WP)**

STANDARD DESCRIPTION

- The WP Volatile Solids standard is provided in a vial that contains approximately 21 mL of concentrate.
- The WP Volatile Solids standard is preserved with nitric acid.
- Store the standard in the unopened vial at room temperature (~ 20 - 25°C).

ADDITIONAL INFORMATION

- The standard has been provided as a concentrate that must be diluted prior to analysis.
- After dilution the standard should be analyzed per your routine procedures for Volatile Solids.

STANDARD PREPARATION, ANALYSIS and STORAGE

1. For best results, the PT standards should be stored at room temperature (~ 20 - 25°C).
2. Add approximately 900 mL of ASTM Type 1 water to a 1000 mL class A volumetric flask.
3. Transfer exactly 10.0 mL of the PT standard concentrate to the flask using a class A volumetric pipette.
4. Bring the flask to volume with ASTM Type 1 water.
5. Mix the solution by inverting the volumetric flask a minimum of three times.
6. The standard is now ready for preparation and analysis per the selected method(s).
7. The sample should be analyzed as soon as possible after dilution,
8. Report all results in mg/L per the reporting instructions contained in this booklet.
9. Store the diluted standard and any remaining concentrate at room temperature (~ 20 - 25°C).

CONCENTRATION RANGE and PTRL

- After preparation per these instructions, the standard will contain Volatile Solids at a certified Concentration within the range shown.
- The NELAC Proficiency Testing Reporting Limit (PTRL) is provided as guidance when analyzing NELAC PT standards. At a minimum, the laboratory should use a method that is sensitive enough to generate quantitative results at the PTRL shown.

Analyte	Units	Concentration Range	PTRL
NELAC Experimental Analytes			
Volatile Solids	mg/L	100 - 500	56.0