

Instructions for: WP Volatile Suspended Solids (Cat # PT-VSS-WP)

SCOPE AND APPLICATION

Wibby Environmental's WP Volatile Suspended Solids Proficiency Testing Standard is designed to be used with promulgated EPA methods as well as applicable methods from consensus organizations such as AWWA and ASTM. If you have any questions about the use of these standards, please contact Wibby Environmental Customer Service at 1-866-WibbyPT (866-942-2978).

SAMPLE PREPARATION AND ANALYSIS

1. For best results, the PT standard should be stored near room temperature (~ 20°C).
2. The standard is ready for preparation and analysis per the selected method(s) as received. No dilutions are required prior to use.

REPORTING RESULTS

1. Report results to three significant figures.
2. Report your results on line at www.wibby.com. Click on the "Online Data Entry" link or the "PT Manage" link.
3. You may also report your results using the Data Reporting Sheets enclosed with your standards. FAX your results to Wibby Environmental at 1-866-283-0269 or mail the results to Wibby Environmental, 6390 Joyce Drive, #100, Golden, CO, 80403.
4. Wibby Environmental must receive all results prior to the study closing date shown on the Data Reporting Sheets.

SAFETY

The standards are designed for use by laboratory professionals who are familiar with handling environmental reference materials as well as hazardous materials. If you have any questions about the safe handling of these standards or require a Material Safety Data Sheet (MSDS), please contact Wibby Environmental at 1-866-WibbyPT (866-942-2978).

QUESTIONS?

If you have any questions regarding these standards or reporting requirements, please call Wibby Environmental at 1-866-WibbyPT (866-942-2978).

6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.wibby.com

Water Pollution Proficiency Testing Concentration Ranges and PTRLs

Definitions:

PTRL

NELAC Proficiency Testing Reporting Limits (PTRLs) are provided as guidance to laboratories analyzing NELAC PT samples. At a minimum, the laboratory should use a method that is sensitive enough to generate quantitative results at the PTRLs shown. (REF: NELAC PT FOT Tables)

NA

Not Applicable (NA) has been applied to analytes where a PTRL is not applicable and to state specific analytes that have not had a PTRL determined by the applicable accrediting agency.

Volatile Suspended Solids (PT-VSS-WP)

NELAC Code	Analyte	Units	Concentration Range	PTRL
Additional State Specific Analytes				
2070	Volatile suspended solids	mg/L	25.0 - 200	NA